



May 21st, 11:45 AM - 12:15 PM

# Low Energy Defibrillation by Synchronization; 90 % less energy compared to one shock.

Flavio H. Fenton

*Georgia Institute of Technology*, [flavio.fenton@physics.gatech.edu](mailto:flavio.fenton@physics.gatech.edu)

Yanyan Ji


*Georgia Institute of Technology*, [yji47@gatech.edu](mailto:yji47@gatech.edu)

Ilija Uzelac

*Georgia Institute of Technology*, [ilija.uzelac@physics.gatech.edu](mailto:ilija.uzelac@physics.gatech.edu)

*See next page for additional authors*

Follow this and additional works at: <http://scholarscompass.vcu.edu/bamm>

 Part of the [Cardiology Commons](#), [Cardiovascular Diseases Commons](#), [Disease Modeling Commons](#), [Dynamic Systems Commons](#), [Medical Biophysics Commons](#), [Non-linear Dynamics Commons](#), and the [Partial Differential Equations Commons](#)

---

<http://scholarscompass.vcu.edu/bamm/2016/May21/6>

This Event is brought to you for free and open access by the Dept. of Mathematics and Applied Mathematics at VCU Scholars Compass. It has been accepted for inclusion in Biology and Medicine Through Mathematics Conference by an authorized administrator of VCU Scholars Compass. For more information, please contact [libcompass@vcu.edu](mailto:libcompass@vcu.edu).

---

**Presenter Information**

Flavio H. Fenton, Yanyan Ji, Ilija Uzelac, Niels Otani, and Elizabeth M. Cherry

By establishing a relationship between the response of cardiac tissue to an electric field and the spatial distribution of heterogeneities due to the coronary vascular structure, it is possible to show that these heterogeneities serve as nucleation sites for the generation of intramural electrical waves with density  $\rho(E)$  in response to a pulsed electric field of strength  $E$ . By using this relationship it is possible to entrain fibrillatory cardiac tissue to a given pacing frequency by synchronization thus terminating the spatiotemporal complex dynamics and restoring the tissue to a normal rhythm. We demonstrate this mechanism first mathematically and then in vitro using voltage optical mapping in atrial and ventricular tissue. We show that the energy required per pulse is about 10% the energy required for a defibrillation shock.